

The definitive version is available at <http://www.iospress.nl/>.

Franck P. Vidal, Pierre-Frédéric Villard, Richard Holbrey, Nigel W. John, Andrew Bulpitt, Derek A. Gould: Developing An Immersive Ultrasound Guided Needle Puncture Simulator. In *Proceeding of Medicine Meets Virtual Reality 17 - NextMed (MMVR17)*. Volume 142 of Studies in Health Technology and Informatics, pages 398-400, Long Beach, California, January 2009. IOS Press.

PMID: 19377193

**Keywords:** image guided needle puncture training, interventional radiology training, needle puncture.

```
@inproceedings{Vidal2009MMVR,
  author = {F. P. Vidal and {P.-F.} Villard and R. Holbrey and N. W. John and
F. Bello and A. Bulpitt and D. A. Gould},
  title = {Developing An Immersive Ultrasound Guided Needle Puncture Simulator},
  booktitle = {Proceeding of Medicine Meets Virtual Reality 17 (MMVR17)},
  year = 2009,
  series = {Studies in Health Technology and Informatics},
  volume = 142,
  pages = {398-400},
  month = jan,
  address = {Long Beach, California},
  annotation = {Jan-19--22, 2012},
  abstract = {We present an integrated system for training ultrasound guided
needle puncture. Our aim is to provide a cost effective and validated
training tool that uses actual patient data to enable interventional
radiology trainees to learn how to carry out image-guided needle
puncture. The input data required is a computed tomography scan of
the patient that is used to create the patient specific models. Force
measurements have been made on real tissue and the resulting data is
incorporated into the simulator. Respiration and soft tissue
deformations are also carried out to further improve the fidelity of
the simulator.},
  keywords = {image guided needle puncture training, interventional radiology
training, needle puncture},
  pmid = {19377193},
  publisher = {IOS Press}
}
```

# Developing An Immersive Ultrasound Guided Needle Puncture Simulator

Franck P. VIDAL<sup>1</sup>, Pierre-Frédéric VILLARD<sup>2</sup>, Richard HOLBREY<sup>3</sup>, Nigel W. JOHN<sup>1</sup>,  
Andrew BULPITT<sup>3</sup>, Derek A. GOULD<sup>4</sup>

<sup>1</sup> School of Computer Science, Bangor University, UK

<sup>2</sup> Biosurgery and Surgical Technology Department, Imperial College, London, UK

<sup>3</sup> School of Computing, The University of Leeds, UK

<sup>4</sup> Royal Liverpool University Hospital, UK

## Abstract


We present an integrated system for training ultrasound guided needle puncture. Our aim is to provide a cost effective and validated training tool that uses actual patient data to enable interventional radiology trainees to learn how to carry out image-guided needle puncture. The input data required is a computed tomography scan of the patient that is used to create the patient specific models. Force measurements have been made on real tissue and the resulting data is incorporated into the simulator. Respiration and soft tissue deformations are also carried out to further improve the fidelity of the simulator.

**Keywords:** image guided needle puncture training, interventional radiology training, needle puncture.

## 1 Introduction

Interventional radiology (IR) uses imaging to guide needles, catheters (tubes) and wires through organ systems using a small portal of entry (minimal access) into the body. As a result, IR techniques generally have less risk, less post-operative pain and shorter recovery time compared to open surgery. Targeted treatments are performed using imaging for guidance in all organ systems, including the liver, the urinary tract, biliary system, gut and blood vessels in the field of many different medical and surgical specialities. Fluoroscopic imaging, computed tomography (CT) and ultrasound (US) are the most commonly used imaging modalities for guidance. US is particularly commonly used for biopsy and nephrostomy [4, 5].

Interventional radiology basic skills are learnt in an apprenticeship in patients, though this is associated with the pain, risk and procedure time penalties of the trainee's inexperienced manipulations. High fidelity simulations using VE offer an alternative, though training of needle puncture US guided by imaging has been largely disregarded. To our knowledge, the Haptic Operative Realistic Ultrasonography Simulator (HORUS) is the only other computer-based simulator to address this problem [1] and is particularly aimed at US-guided needle insertion for biopsy and radio frequency thermal ablation. Our aim is to address this deficiency by providing a cost effective, configurable and validated training tool to enable IR trainees to learn how to carry out image-guided needle puncture. Our initial simulation platform made use of static CT volume data without co-location between the haptic devices and the 3D virtual environment [9]. Ideally, imaging data sets for simulation of image-guided procedures would alter dynamically in response to deformation forces such as respiration and needle insertion. We describe a methodology for deriving such dynamic



**Figure 1:** Design and implementation of the simulator.

volume rendering from patient imaging data. The VE is delivered using stereoscopy with co-location between the haptic devices and the 3D virtual environment to provide a realistic immersive simulation.

## 2 Tools and Methods

For this project, US guided nephrostomy and US guided needle biopsy of the liver will be used to illustrate the validity of our hypothesis (see Figure 1(a)).

With patient consent, selected, routine imaging (computed tomography, magnetic resonance, ultrasound) of straightforward and complex anatomy and pathology was initially anonymized and uploaded to a repository. Data sets are then segmented using interactive processes to label target anatomy. This is used to create surface (triangular) meshes and volume (tetrahedral) meshes (see Figure 1(b)).

We now focus on the mechanical behaviour. Firstly the liver undergoes deformations due to the needle insertion. The detection collision between the needle and the liver is performed using H3D [6]. This API allows us to know the area of the collision and the magnitude of the touching force. The computation of this mechanical behaviour is then done using SOFA [7]. The tetrahedron mesh is approximately made of 500 nodes per liver. The deformation is then computed in real-time using the mass-spring method. Secondly, the liver has its own intrinsic motion and deformation due to the respiration. It is attached to the diaphragm with ligaments which follow this movement. We then developed our own diaphragm model based on rib kinematics and muscle contraction/relaxation [10].

The force feedback delivered in virtual environment simulators is generally an approximation to a real procedure, as assessed by experts. Haptics based on actual procedural forces should allow a more realistic simulation of the subtle cues [2]. In evaluating needle puncture procedures, for example, *in vitro* studies are essential for detailed understanding of the physical components and effects of overall summated forces and we are collecting such experimental data [8].

As the trainee moves the image probe over the skin of the virtual patient, an US-like image corresponding to the position and orientation of the image plane is computed and displayed in real-time (see Figure 1(c)).

## 3 Results

An integrated simulator for image guided needle puncture is being produced. Data sets were derived from 10 patients and converted into deformable VEs.

The simulator has been developed in close collaboration with practicing interventional radiologists. In particular, we have made use of IR task analysis studies that have been carried out by our collaborators [3]. Preliminary content validation studies of a framework developed for training on

imaging guided nephrostomy and liver biopsy procedures, demonstrated favourable observations that are leading to further revisions, including implementation of an immersive VE.

## 4 Conclusion

Anonymised patient datasets form an essential basis for construction of ‘virtual reality’ training models. Such models can be developed and validated in the academic environment. Initial results predict well for the fidelity of this training simulator. Finally, comprehensive construct and concurrent validation, and skills transfer studies, form a key part of future, proposed work. This will be conducted within IR curricula to evaluate the clinical effectiveness of the simulator to train visceral needle puncture procedures.

## References

- [1] C. Forest *et al.* Ultrasound and needle insertion simulators built on real patient-based data. In *Proceedings of MMVR 15*, pages 136–139, 2007.
- [2] O. Gerovichev *et al.* The effect of visual and haptic feedback on computer-assisted needle insertion. *Computer Aided Surgery*, 6(9):243–249, 2004.
- [3] S. J. Johnson *et al.* Physical and cognitive task analysis in interventional radiology. *Journal of Clinical Radiology*, 1(61):97–103, 2006.
- [4] T. A. Matalon and B. Silver. US guidance of interventional procedures. *Radiology*, 1(174):43–47, 1990.
- [5] J. P. Pedersen. Percutaneous nephrostomy guided by ultrasound. *Journal of Urology*, 2(112):157–159, 1974.
- [6] SenseGraphics. H3D. <http://www.h3d.org/>. (last visited October 2008).
- [7] SOFA Project. Simulation open framework architecture (SOFA). <http://www.sofa-framework.org/>. (last visited October 2008).
- [8] F. P. Vidal *et al.* Force penetration of chiba needles for haptic rendering in ultrasound guided needle puncture training simulator. In *MICCAI 2008 Workshop “Needle Steering: Recent Results and Future Opportunities”*, pages 136–139, 2008. Available online at <http://lcsr.jhu.edu/NeedleSteering/Workshop/Vidal.html> (last visited October 2008).
- [9] F. P. Vidal *et al.* Simulation of ultrasound guided needle puncture using patient specific data with 3D textures and volume haptics. *Computer Animation and Virtual Worlds*, 19(2):111–127, 2008.
- [10] P.-F. Villard *et al.* Modelling organ deformation using mass-springs and tensional integrity. In *Proceedings of the International Symposium on Computational Models for Biomedical Simulation 2008 (ISMBS ’08)*, pages 221–226, 2008.